

Automatic lateral calibration of tunneling microscope scanners

Rostislav V. Lapshin^{a)}

Zelenograd Physical Problems Institute, Zelenograd, Moscow, 103460, Russia

(Received 6 January 1998; accepted for publication 6 May 1998)

A practical method is described to find automatically the calibration coefficients and residual nonorthogonality of a tunneling microscope scanner. As initial data, the coordinates of three atoms were used forming a triangle in a highly oriented pyrolytic graphite surface appearing in the form of a spatially geometrical measure. A recognition procedure is described which can be applied to determine the lateral coordinates of the atoms. Length and orientation distortions were calculated, estimates of calibration errors were given and the requirement on the nonorthogonality limit was formulated for manipulator a given that ensures measurements of the predetermined accuracy. The sensitivity of the method to a noise in atom coordinates was determined. Experimental data showing the practical suitability of the method developed are presented. © 1998 American Institute of Physics. [S0034-6748(98)00308-6]

I. INTRODUCTION

Today the scanning tunneling microscope (STM) is the ideal instrument for surface investigations of matter on an atomic scale; nevertheless, it possesses certain shortcomings which are mainly various kinds of distortion in obtained data.

It is known, for example, that in due course piezoceramics of the STM manipulators are dipoling, which leads to changes in the strain coefficients and errors in measurement. For elimination of the latter it is necessary to recalibrate the microscope periodically. However, manual calibration is a rather laborious process.

The subject of the present article includes questions concerning procedures of automatic determination of calibration coefficients¹ and an oblique angle, analysis of encountered errors, as well as finding the most suitable calibration objects.

This article also can be considered as an introduction to a special scanning method based on local recognition and connection of surface features.¹ The approach permits us to suppress noise and to get rid of image distortions caused by therm drift,²⁻⁵ creep,⁶ and hysteresis^{2,5} practically completely. As a result the image contains only distortions produced by errors in determination of the calibration coefficients and oblique angle^{3,4,7} as well as by size and form of the tip.⁸ Since the tip distortions do not disturb relative positions of the features composing calibration structure, true values of the calibration coefficients and oblique angle can be extracted from the image.

II. DESCRIPTION OF THE METHOD

A. Procedure for the determination of the lateral coordinates of atoms

Often when investigating a surface of solids at the atomic scale, it is sufficient to have only information about

the atom coordinates. Since the locations of the atom centers on the surface image obtained with the help of STM are not defined, a procedure should be developed which permits them to be found. The procedure desired is a recognition program that operates as follows.

The atomic relief of a surface given in a certain window is cut by a horizontal plane at height z . In other words, an operation of image segmentation is being carried out where the image components of interest are separated (i.e., atoms from the background) by threshold z . At that, a set of *a priori* noncrossing plane domains are being formed; at times, they have local violations of convexity, and are of noncircular shape.

The domain differences are brought forth by noise in a tunnel current stabilization system and scanning channels, which gives "torn" edges, by therm drift and creep as well as by uneliminated distortions connected, in general, with geometrical dimensions of the probe tip and its shape. Quantitatively these differences can be evaluated by calculating the domain "compactness:"⁹ $C = 4\pi S/L^2$, where S is the domain square, and L is the domain border length (it is easy to check out that in the case of a circle this magnitude is "perfect," i.e., equal to unity). Thus, by defining the gravity centers of the obtained domains, the lateral coordinates of the atom centers on the surface can be found.

To determine the gravity centers it is necessary to know the coordinates of the points lying inside a domain and on its border, i.e., to be able to find out the domain contour and its closure, as well as to distinguish domains from each other.

The algorithm developed invokes steps similar to those described in the works of Sidhu and Boute,⁹ Freeman,¹⁰ Rosenfeld and Kak,¹¹ and represents the following sequence of operations:

- (1) An image given in a certain window is looked through until a point that has the height greater or equal to z is met. This is the first point belonging to the contour of a regular domain. The point found becomes the current point.

^{a)}Electronic mail: au135@elvis.msk.su; rlapshin@openmail.irex.ru

- (2) Eight points neighboring on the current one and forming the chain code¹⁰ are looked through until a point is found with a height greater or equal to z . The point found that way becomes the current point.
- (3) Item (2) is repeated until the first point of the domain border is met, i.e., the closure of the domain contour occurs.
- (4) Knowing the coordinates of the domain contour, the coordinates of the gravity center of the domain at subpixel level can be calculated.
- (5) Image points belonging to the recognized domain are assigned zero values. Provided the row and the column counters of the window being looked through do not reach the maximum value, a jump is made to item (1).

Note, that item (2) of the algorithm implies that the round trip over the neighbors is made, first, each time in the same direction and, second, in such a way that the initial point of the trip be an external point.

The recognition procedure permits direct calculation of the average lattice constant as well as, beside the atoms, to detect and to run statistics of other objects encountered in the image, such as isolated noise peaks (solitary pixels), isolated noise islets and threads (chains of pixels), merged atomic domains. Those recognized atoms situated near the edge of the image or cut by it are automatically rejected since the data collected on those domains are either much distorted because of the influence of edge effects or just wrong. Note here that long threads appearing parallel to the fast scan direction reveal surface contamination.

The larger squares of the atomic domains under investigation are, the more precisely the atom centers can be determined. The domain sizes depend on the cutting height z . Therefore, making "scanning" of surface relief along z from top to bottom and doing recognition iterations (25...35 times for image with 2^{16} gray levels) it becomes possible to find such a height that being decreased by a minimum step, the atom domains start to merge with each other. The height thus obtained may be accepted as the optimal cut height.

The procedure is written in Pascal, the rate of recognition of atoms with 10...12 pixel contour, with simultaneous classification of the objects mentioned above, and statistics accumulation on a computer like the IBM PC 386DX/33 MHz, makes up 1000 atoms/s, approximately. For application of the procedure the image is recommended to be pre-smoothed, and the trend to be removed from it.

As an illustration to recognition procedure working the stylized image (272 atoms, average lattice constant 2.717 \AA) is shown in Fig. 1(b). This picture is obtained as a result of recognition of midsize STM¹² scan of highly oriented pyrolytic graphite (HOPG) in Fig. 1(a).

The recognition procedure can be considered as a kind of filter-compressor since, at one hand, it "smooths" initial image producing a stylized surface, at other hand, it "squeezes" the data extracting the most valuable information from the picture.




FIG. 1. (a) An example of midsize graphite scan (128×128 pixels, microscope is decalibrated). The measurements were made in constant Z mode at $U_{\text{tun}} = -100 \text{ mV}$, $I_{\text{tun}} = 300 \text{ pA}$. Average compactness 97.2%. (b) The result of recognition-stylized surface of graphite. Carbon atoms are shown as hemispheres.

B. Determination of the calibration coefficients and the angle of obliquity

The idea of the method consists in searching for the parameters for linear transformation of a distorted image with an *a priori* known structure which leads to obtaining the correct image. After that, using the parameters found, the same kind of manipulation is produced but now with images whose structure is unknown.

In order to find three unknowns, i.e., the \bar{K}_x , \bar{K}_y gains along the X , Y axes of the microscope piezomanipulators,

respectively, and the oblique angle α , it is necessary to form a system of three equations. For that, in the general case it is sufficient to know the lateral coordinates of calibration surface six atoms, which form three different segments $|AB| = k_1 a_1$, $|CD| = k_2 a_2$, and $|EF| = k_3 a_3$ (where a_1, a_2, a_3 are lattice constants, k_1, k_2, k_3 are integer coefficients).

Since in an oblique coordinate system Pythagoras's theorem is not valid, for the next constructions the following linear transformation from an oblique system $\bar{x}\bar{o}\bar{y}$ connected with X, Y STM piezomanipulators to the rectangle xoy will be used

$$\begin{aligned} x &= \bar{K}_x \bar{x} + \bar{K}_y \bar{y} \sin \alpha, \\ y &= \bar{K}_y \bar{y} \cos \alpha. \end{aligned} \quad (1)$$

Further, a HOPG surface will be used as the calibration surface where the distances between β atoms¹³ are the same $a_1 = a_2 = a_3 = a$ and equal to 2.464 Å.¹⁴ To fix ideas, a particular case is considered where $k_1 = k_2 = k_3 = k = 2, 4, 8, \dots$, and the measure segments placed so that their center points cross in one point, and the ends form a rectilinear hexagon; this will be named a "star" scheme. Thus, in the method under description, HOPG surface atoms and geometrical structure resulting from them serve as a spatially geometrical standard.

Using expression (1) we solve the system of equations composed for star beams. Finally, we obtain

$$\bar{K}_x = ka \sqrt{|G/H|}, \quad (2)$$

$$\bar{K}_y = ka \sqrt{|I/H|}, \quad (3)$$

$$\alpha = \arcsin \frac{|H| - J^2 |G| - K^2 |I|}{2JK \sqrt{|GI|}}, \quad (4)$$

where $G = y_{CD} y_{EF} (x_{CD} y_{EF} - x_{EF} y_{CD}) - y_{AB} y_{EF} (x_{AB} y_{EF} - x_{EF} y_{AB}) + y_{AB} y_{CD} (x_{AB} y_{CD} - x_{CD} y_{AB})$; $H = (x_{CD} y_{EF} - x_{EF} y_{CD})(x_{AB} y_{EF} - x_{EF} y_{AB})(x_{AB} y_{CD} - x_{CD} y_{AB})$; $I = x_{CD} x_{EF} (x_{CD} y_{EF} - x_{EF} y_{CD}) - x_{AB} x_{EF} (x_{AB} y_{EF} - x_{EF} y_{AB}) + x_{AB} x_{CD} (x_{AB} y_{CD} - x_{CD} y_{AB})$; $J = x_{AB}$; $K = y_{AB}$; the designation x_{AB} means the $x_A - x_B$ difference and $y_{AB} = y_A - y_B$ (the same for the other point pairs).

The last result can be obtained by using the property of equality of star beams. From the property only two independent equations can be formed but there are three variables. If Eq. (3) is divided by Eq. (2), the quantities a in the both coefficients will cancel out and the ratio \bar{K}_y / \bar{K}_x will then not depend on the a . Hence, introducing instead of the \bar{K}_x, \bar{K}_y coefficients the reduction coefficients $\bar{K}_x^* = 1, \bar{K}_y^* = \bar{K}_y / \bar{K}_x$, and solving the system with respect to the variables \bar{K}_y^*, α we obtain

$$\bar{K}_y^* = \sqrt{|I/G|}, \quad (5)$$

$$\alpha = \arcsin \frac{L|G| + M|I|}{2N \sqrt{|GI|}}, \quad (6)$$

where $L = x_{AB}^2 - x_{CD}^2$; $M = y_{AB}^2 - y_{CD}^2$; $N = x_{CD} y_{CD} - x_{AB} y_{AB}$.

Thus, using the reduction calibration coefficients it is sufficient to carry out the calibration of one microscope manipulator in the usual way; the calibration coefficient of another manipulator can be made from their ratio, first having calculated a value of the reduction coefficient by the means of formula (5).

Let us now consider another particular case when the initial calibration structure is an equilateral triangle ($k=1, 2, 3, \dots$). It is easy to see that the formulae for calculation of the calibration coefficients and the oblique angle by "triangle" scheme have the form of Eqs. (2)–(4), (5), (6) deduced earlier for the star; where instead of the former CD, EF indexes, the BC, AC ought to be used, accordingly.

C. Character and distortion values of a nonorthogonal STM scanner

In practice, in measurements carried out by STM images, the lattice constants and the angles between directions have most frequently to be determined. Therefore, if in the capacity of a characteristic element of the image a segment is chosen, its distortions caused by the scanner nonorthogonality can be subdivided into the distortions connected with the change of its length Δl and orientation Δo .

Using polar coordinates ($x = l_0 \cos \beta_0, y = l_0 \sin \beta_0$), and also assuming $\bar{K}_x = \bar{K}_y = 1$, the distortions can be presented as follows:

$$\delta l = \frac{\Delta l}{l_0} 100\% = \left| \frac{1}{\cos \alpha} \sqrt{\cos^2(\alpha + \beta_0) + \sin^2 \beta_0} - 1 \right| 100\%, \quad (7)$$

$$\Delta o = \left| \arctan \frac{\sin \beta_0 [\cos(\alpha + \beta_0) - \cos \beta_0]}{\cos \beta_0 \cos(\alpha + \beta_0) + \sin^2 \beta_0} \right|, \quad (8)$$

where l_0, β_0 are length and orientation (counted from ox axis in counter clockwise direction) of an initial nondistorted segment, accordingly. It is obvious from the above formulas that, if a segment under investigation is parallel to some coordinate axis, that is, $\beta_0 = 0$ or $\beta_0 = \pi/2 - \alpha$, then the distortions $\delta l = 0, \Delta o = 0$ or $\delta l = 0, \Delta o = \alpha$, accordingly.

Thus, the distortions of a nonorthogonal scanner are such that the segments placed parallel to the axes of the manipulators or to the coordinate axes on the screen do not suffer length distortions. Therefore, the commonly used method for microscope calibration by orienting the segment to be measured parallel to the axes works properly, though in practice it is not very suitable as it requires a sample rotation. It should be taken into consideration that the segments oriented in other directions are distorted, the degree of the distortions depending on the direction.

Let us determine the direction along which the segment is distorted most. For this, it will be rotated about the origin




FIG. 2. The dependencies on obliquity angle α of maxima distortions of segment length δl and orientation Δo , as well as relative error δl_{clb} of measurement of linear dimensions when attention to nonorthogonality scanner at calibration is not paid ($\beta_0 = \beta_0^{max}$).

of coordinates and that direction will be sought where δl and Δo are maxima. Differentiating Eqs. (7) and (8) by the variable β_0 yields that the maximum length distortion of segment appears when it is oriented at an angle of

$$\beta_0^{max} = \arctan \frac{1 \pm \sin \alpha}{\mp \cos \alpha} \quad (9)$$

for $\pm \alpha$, accordingly, and the maximum orientation distortion

$$\beta_0^{max} = \pm \frac{\pi}{2} - \frac{3\alpha}{4}. \quad (10)$$

The dependencies $\delta l(\alpha)$ and $\Delta o(\alpha)$ which are shown in Fig. 2 are derived by substituting, respectively, $\beta_0 = \beta_0^{max}$ into formulas (7) and (8).

D. Calibration object for direct determination of the calibration coefficients and angle of obliquity

Let us find now eigenvalues λ and eigenvectors of linear transformation inverse to Eq. (1). For that purpose a characteristic equation is composed

$$\begin{vmatrix} \frac{1}{\bar{K}_x} - \lambda & \frac{-\tan \alpha}{\bar{K}_x} \\ 0 & \frac{1}{\bar{K}_y \cos \alpha} - \lambda \end{vmatrix} = 0, \quad (11)$$

solving which, one readily find that $\lambda_1 = 1/(\bar{K}_y \cos \alpha)$, $\lambda_2 = 1/\bar{K}_x$. Substituting the eigenvalues obtained into system of equations

$$\begin{aligned} (1 - \lambda \bar{K}_x)x - \tan \alpha y &= 0, \\ (1 - \lambda \bar{K}_y \cos \alpha)y &= 0, \end{aligned} \quad (12)$$

and solving, we deduce that a set of segments placed along abscissa axes $\beta = \beta_0 = 0$, as well as at an angle of $\beta = \beta_0 = -\alpha/2$ ($\bar{K}_x = \bar{K}_y$) represents a family of eigenvectors of the linear transformation inverse to Eq. (1). Therefore, the orientation of the vectors in the transformation remains undistorted.

Thus, the above analysis shows that the calibration object for direct measurements of calibration coefficients and an oblique angle must possess circular symmetry for excluding the necessity of orientation of the calibration segment along axes of coordinates and has to have a mark indicating its center. A good candidate here is a ring or series of concentric rings with known radii and height (for calibration of the Z manipulator).

The calibration procedure by the object under discussion can be reduced to the following steps.

- (1) Measure the lengths of two segments formed by the intersection of straight lines parallel to coordinate axes on the screen and passing through the center mark with the ring. The segments give calibration coefficients \bar{K}_x and \bar{K}_y .
- (2) Transform the structural image to obtain scales along coordinates axes equal to each other ($\bar{K}_x = \bar{K}_y$).
- (3) Search such a direction of a straight line going through the structure center where the segment length formed at intersection of this line with the ring is equal to $\lambda_1 D$ (D is the ring diameter). The direction obtained in that way corresponds to half the value of an obliquity angle being sought.

Note, that the method of direct calibration suggested here can be easily automated by applying the recognition procedure described in Sec. II A; for this, a mark of the structure center is no longer required.

E. Errors of STM calibration made without taking into account nonorthogonality of the scanner. Formulation of the requirement to limit allowable nonorthogonality in the manufacture of manipulators

Let us estimate the value of systematic calibration error which influences a measurement error of linear dimensions in a case when nonorthogonality of STM manipulators is neglected. For this, two pairs of calibration coefficients calculated from a skewed image must be found. One pair is K_x and K_y which corresponds to coefficients obtained as if the microscope scanner is orthogonal, i.e., without attention to its obliquity; and the second— \bar{K}_x , \bar{K}_y , vice versa, with regard for it. For this, the latter must be expressed not only in differences of initial coordinates but also through the angle α which represents in that case its parameter. Finally, we obtain

$$\bar{K}_x = k\alpha \sqrt{\frac{-MO - 2KR \sin \alpha (N \sin \alpha \pm \sqrt{N^2 \sin^2 \alpha - LM})}{P(O^2 - 4JKQR \sin^2 \alpha)}}, \quad (13)$$

$$\bar{K}_y = k\alpha \sqrt{\frac{LO + 2JQ \sin \alpha (N \sin \alpha \mp \sqrt{N^2 \sin^2 \alpha - LM})}{P(O^2 - 4JKQR \sin^2 \alpha)}}, \quad (14)$$

where $O = x_{AB}y_{CD} + x_{CD}y_{AB}$, $P = x_{AB}y_{CD} - x_{CD}y_{AB}$, $Q = x_{CD}$, $R = y_{CD}$; and a pair of K_x , K_y coefficients appears from the expressions above at the passage to the limit

$$K_x = \lim_{\alpha \rightarrow 0} \bar{K}_x = k\alpha \sqrt{|M/(OP)|}, \quad (15)$$

$$K_y = \lim_{\alpha \rightarrow 0} \bar{K}_y = k\alpha \sqrt{|L/(OP)|}, \quad (16)$$

in fact, the coefficients are calibrations of the microscope with an orthogonal scanner obtained by two measured segments. Note, that by their meaning both pairs of coefficients must be positive, nonzero real numbers.

The maximum relative error of length measurement of some segment placed at an angle of Eq. (9), i.e., at an angle when its distortions manifest themselves most strongly, can be found by the formula

$$\begin{aligned} \delta l_{\text{clb}} &= \frac{\Delta l_{\text{clb}}}{l_0} 100\% \\ &= \left| \frac{1}{\bar{K}_x \bar{K}_y \cos \alpha} \right. \\ &\quad \times \sqrt{K_x^2 \bar{K}_y^2 \cos^2(\alpha + \beta_0^{\text{max}}) + K_y^2 \bar{K}_x^2 \sin^2 \beta_0^{\text{max}}} \\ &\quad \left. - 1 \right| 100\%. \end{aligned} \quad (17)$$

Figure 2 shows the character of change of δl_{clb} as a function of an obliquity angle α .

Let us solve an inverse problem connected with searching for such an obliquity angle, which leads to a given error magnitude. It can be concluded from the $\delta l(\alpha)$ graph that for achieving measurements of linear dimensions of surface elements with a relative error not worse than 1%, for instance, the technology of STM scanner must guarantee the nonorthogonality value of the manipulator's axes within the limits of the range $\pm 1.13^\circ$. At the same time, from the $\delta l_{\text{clb}}(\alpha)$ curve it is easy to see that manipulator nonorthogonality may be neglected agreeing to an error, for example, in 1% if α by absolute value does not exceed 0.29° .

F. Sensitivity of the calibration coefficients and the oblique angle to errors in coordinates of initial data. Comparison of the triangle and star schemes

Let us define the method's sensitivity to errors in measurement of the atom positions, which are taken as the initial data for triangle and star schemes. As the calibration coefficients and the oblique angle are not being determined by a direct measurement, but rather computed by formulas, for estimation of errors (root mean squares) contained in their values the following expression must be calculated:

$$\sigma_f = \sqrt{\left(\frac{\partial f}{\partial x_A}\right)^2 \sigma_{x_A}^2 + \left(\frac{\partial f}{\partial x_B}\right)^2 \sigma_{x_B}^2 + \dots + \left(\frac{\partial f}{\partial x_F}\right)^2 \sigma_{x_F}^2 + \left(\frac{\partial f}{\partial y_A}\right)^2 \sigma_{y_A}^2 + \left(\frac{\partial f}{\partial y_B}\right)^2 \sigma_{y_B}^2 + \dots + \left(\frac{\partial f}{\partial y_F}\right)^2 \sigma_{y_F}^2}, \quad (18)$$

where the atom centers' coordinates x_A, \dots, x_F , y_A, \dots, y_F measured with root mean square errors $\sigma_{x_A}, \dots, \sigma_{x_F}$, $\sigma_{y_A}, \dots, \sigma_{y_F}$, respectively, are put as variables, and the designation f , while calculating, is substituted by the functions: $\bar{K}_x, \bar{K}_y, \alpha$.

To simplify, let us assume that all errors in the atom center coordinates are equal to σ . For numerical estimates the σ value will be accepted less, at least by order,² than lateral resolution of the microscope, which amounts to $\sim 1 \text{ \AA}$. It corresponds to the relative error of measurement of graphite lattice constant length and makes up $\sim 4\%$.

In view of what was said above, the formula (18) can be written as

$$\sigma_f = \sigma \sqrt{\sum_{i=A,B,\dots,F} \left[\frac{1}{\bar{K}_x^2} \left(\frac{\partial f}{\partial x_i}\right)^2 + \frac{1}{\bar{K}_y^2} \left(\frac{\partial f}{\partial y_i}\right)^2 \right]}. \quad (19)$$

The multipliers before the partial derivatives are normalizing factors; they eliminate the error value dependence upon the particular decalibration degree of the microscope. It is necessary to note that the error values $\sigma_{\bar{K}_x}$, $\sigma_{\bar{K}_y}$, and σ_α are inversely proportional to the measure length $k\alpha$, therefore, it is reasonable to use calibration structures with large values of factor k .

Since the errors $\sigma_{\bar{K}_x}, \sigma_{\bar{K}_y}$ themselves bring little information for the analysis of noise influence, it would be useful to pass over to error σ_l of segment length determination (where

$l(x_i, y_i) = \sqrt{x^2(x_i, y_i) + y^2(x_i, y_i)}$, which is to be found as it was done previously.

Quantities $x(x_i, y_i)$, $y(x_i, y_i)$ are calculated as follows. First, using transformation inverse to Eq. (1), let us express the coordinate differences \bar{x}, \bar{y} of some segment l via coordinates x, y assuming \bar{K}_x, \bar{K}_y , and α as constants. After that,

$$x(x_i, y_i) = \frac{l_0 \{ \bar{K}_x(x_i, y_i) \bar{K}_y \cos[\alpha + \beta_0^{\max}(\alpha)] + \bar{K}_x \bar{K}_y(x_i, y_i) \sin \alpha(x_i, y_i) \sin \beta_0^{\max}(\alpha) \}}{\bar{K}_x \bar{K}_y \cos \alpha},$$

$$y(x_i, y_i) = \frac{l_0 \bar{K}_y(x_i, y_i) \cos \alpha(x_i, y_i) \sin \beta_0^{\max}(\alpha)}{\bar{K}_y \cos \alpha}, \tag{20}$$

where $i=A, B, \dots, F$.

Formula (19) is intended for computing the methodical errors of the star scheme. By omitting terms with D, E, F indexes and using the appropriate expressions for \bar{K}_x, \bar{K}_y , and α , it is possible to get the formula for definition of the errors in the triangle scheme. Since expression (19), when resolved analytically, looks bulky, the errors were determined numerically.

The obtained results can be analyzed by means of Fig. 3, which shows: the root mean square error σ_α of determination of an obliquity angle and relative error $\delta l_{\text{sys}} = \sigma_l / l_0 \cdot 100\%$ of determination of segment length produced by systematic errors of calculation of calibration coefficients and obliquity angle. The errors shown in the figure were obtained for two different schemes of calculations—triangle and star.

Analysis of the graphs leads to the conclusion that the oblique angle, determined by the triangle scheme, has a systematic error slightly less than the one obtained by the star




FIG. 3. Methodical error σ_α of determination of oblique angle and relative methodical error δl_{sys} of measurement of a linear segment for schemes of triangle and star ($\beta_0 = \beta_0^{\max}$).

apply transformation (1) where calibration coefficients and obliquity angle are now functions of coordinates x_i, y_i , i.e., $\bar{K}_x(x_i, y_i), \bar{K}_y(x_i, y_i), \alpha(x_i, y_i)$. By passing to polar coordinates, finally obtain

scheme. As for the errors of segment length determination, there is practically no difference detectable between the schemes. The presented graphs show that at the error value $\sigma = 0.1 \text{ \AA}$ ($k=2, -15^\circ \leq \alpha \leq +15^\circ$) the methodical error σ_α (triangle scheme) lies in the range of $2.24^\circ - 2.33^\circ$. Therefore, the method offered does not provide a precise determination of small oblique angles under high level of noise in the microscope. For example, at the error σ in determination of the atom center as low as $\sim 0.03 \text{ \AA}$, the relative error of measurement of a linear segment does not exceed 1% and the error of an oblique angle— 1° .

III. EXPERIMENTAL RESULTS

A. Automatic calibration of the microscope and obliquity correction of the graphite image

Figure 4(a) shows a distorted STM¹⁵ image of HOPG surface undergone smoothing, with trend removed. By using the recognition procedure the lateral coordinates of the carbon atoms were determined [see Fig. 4(b), a short summary of recognition results is added] and a surface image was built stylized by hemispheres [see Fig. 4(c)]. With triangles $A_1 B_1 C_1$ and $A_2 B_2 C_2$ using formulas (2), (3), (6) the calibration coefficients $\bar{K}_x = 0.965, \bar{K}_y = 0.940$, and the oblique angle $\alpha = 9.037^\circ$ ($k=4$) were calculated and averaged. Figures 4(d) and 4(e) demonstrate the corrected images (average lattice constant is equal to 2.473 \AA , $\delta\alpha = 0.4\%$). The first picture was obtained by applying a lineshifting algorithm to the image in Fig. 4(a), and the second one, by directly applying formula (1) to the data shown in Fig. 4(b). Coordinates z of atoms in all stylized images are reduced to monolayer (i.e., they are supposed to be equal).

B. Examples of application of recognition procedure

In Fig. 5(a) an atomic force microscope¹² (AFM) image (tapping mode) of nanostructured aluminum surface¹⁶ which was formed by electrochemical polishing of textured aluminum foil is shown. Using recognition procedure [see Fig. 5(b)] one can find number of elements, their average diameter (884 \AA), average spacing between features (2034 \AA),




FIG. 4. (a) The initial distorted image of graphite surface (constant Z mode, $U_{\text{tun}} = 100$ mV, $I_{\text{tun}} = 400$ pA). (b) The auxiliary intermediate image of the atom domains appeared after drawing the horizontal cutting plane at an optimal height. Domains which was recognized as atoms has “+” mark of probable center. (c) Stylized surface image (atoms placed along image edges were discarded). (d) The surface image after correction by lineshifting algorithm. (e) The stylized surface image after correction by direct application of transformation (1).




FIG. 4. (continued).

average height (41 Å). In addition variances and histograms of distributions for these magnitudes can be built.

The method described has been employed also for studying properties of atomically smooth films: MePhSiCl₂, VOCl₃, TiCl₄, which are themselves ordered structures¹⁷ made by the method of chemical assemblage. As the substrate, HOPG and mica (muscovite) were used serving as a conducting and dielectric matrix, respectively. With this, the




FIG. 5. (a) AFM scan of aluminum surface which was made at electrochemical polishing process. (b) Plane domains belonging to recognized surface elements are shown in lighter color.

matrix surface was accessible because of the discontinuity of the thin (30 Å) film. Since the research was carried out in air, where trustworthy spectroscopic data can hardly ever be obtained, the “rightness” of the surface topography has been of extreme significance for analysis of the structures under investigation. That is why a great attention was dedicated to revealing and compensating systematic errors as well as to searching for reliable calibration procedures for the instrument.

IV. DISCUSSION

The recognition procedure described is used for measurement of lattice constants, determination of the calibration coefficients and the manipulators’ nonorthogonality. The procedure is useful because, first, it excludes a subjective factor from the measuring process, and second, saves much memory (15–20 times) and processing time while picture

generation since now, instead of a surface image array, it is sufficient to store the coordinates of atom centers, displaying the image of atoms in a stylized form.

The triangle scheme proved to be a preferred solution compared with the star scheme since the former requires twice fewer initial data for realization as the latter with comparable methodical errors. It is also necessary to keep in mind that the calculated values of calibration coefficients and oblique angle are noise sensitive. This circumstance limits the precision achieved by the method in the determination of those quantities.

The compactness value averaged by several domains is a useful numerical characteristic that shows the microscope noise value and/or points out to the presence of local distortions.

It is important that the process of microscope calibration can be readily automated. By the way, recognition of the atom center within some set of calibration structures necessary for averaging of calibration coefficients and the obliquity angle could be substituted by determination of maxima in the space frequency spectrum calculated by a fast Fourier transform algorithm. After that, it becomes possible to extract the parameters of affine transformation directly from the spectrum as was suggested in works of J. F. Jørgensen *et al.*⁴ and S. Carrara *et al.*⁷ Choice of the best method here depends on the particular task under investigation.

As a rule, since different sources lead to the same sort of distortion, it is hard to point out their real origin or separate the influence of one source from another. Therefore, provided the surface image is subjected to the influence of a whole “bouquet” of systematic distortion factors, the method described permits one to determine the calibration coefficients and the oblique angle of some hypothetical scanner. Then, the correction procedure will help partially correct the image by deleting the linear parts of distortions caused by all those factors.

ACKNOWLEDGMENTS

This work was supported by Russian Foundation of Fundamental Researches (project 96-02-17870) and Fund of Physics of Solid Nanostructures (project 96-2012). The author thanks A. V. Denisov for providing the microscope, A. V. Emelyanov and V. N. Ryabokon for organization of work and support, V. V. Efremov and O. E. Lyapin for checking the manuscript and all those mentioned above for advice and suggestions said during discussions.

¹H. Kawakatsu and H. Kougami, *J. Vac. Sci. Technol. B* **14**, 11 (1996).

²D. W. Pohl, *IBM J. Res. Dev.* **30**, 417 (1986).

³V. Yu. Yurov and A. N. Klimov, *Rev. Sci. Instrum.* **65**, 1551 (1994).

⁴J. F. Jørgensen, L. L. Madsen, J. Garnæs, and K. Carneiro, *J. Vac. Sci. Technol. B* **12**, 1698 (1994).

⁵R. V. Lapshin, *Rev. Sci. Instrum.* **66**, 4718 (1995).

⁶S. Vieira, *IBM J. Res. Dev.* **30**, 553 (1986); E. P. Stoll, *Ultramicroscopy* **42–44**, 1585 (1992).

⁷S. Carrara, P. Facci, and C. Nicolini, *Rev. Sci. Instrum.* **65**, 2860 (1994).

⁸E. J. Snyder, E. A. Eklund, and R. S. Williams, *Surf. Sci. Lett.* **239**, L487 (1990).

⁹G. S. Sidhu and R. T. Boute, *IEEE Trans. Comput.* **C-21**, 1206 (1972).

¹⁰H. Freeman, *IRE Trans. Electron. Comput.* **EC-10**, 260 (1961).

¹¹A. Rosenfeld and A. C. Kak, *Digital Picture Processing* (Academic, New York, 1976).

¹²Microscope Solver P4 from NT-MDTC Co.

¹³D. Tománek, S. G. Louie, H. J. Mamin, D. W. Abraham, R. E. Thomson, E. Ganz, and J. Clarke, *Phys. Rev. B* **35**, 7790 (1987).

¹⁴P. Trucano and R. Chen, *Nature* **258**, 136 (1975).

¹⁵The STM of the Physical Chemistry Institute of Russian Academy of Sciences.

¹⁶Sample was made by S. A. Gavrilov (Zelenograd Physical Problems Institute).

¹⁷A. V. Emelyanov and V. V. Egorin, *Journal of Surface: Physics, Chemistry, Mechanics* (in Russian) **1**, 94 (1988); S. D. Dubrovskiy, A. V. Emelyanov, A. V. Zimin, V. N. Inkin, A. A. Malkov, A. A. Malygin, and S. M. Portnov, *J. Appl. Chem.* (in Russian) **65**, 2259 (1992).